
Midmark M11 - v1.0.5
Total Cycles: 2819

Name: RITTER2
Sterilizer ID: V2313228
08/30/2023 01:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.1 F 0.0
2:00 145.5 F 0.0
4:00 176.2 F 0.0
6:00 198.9 F 3.1
8:00 218.6 F 7.2
10:00 231.1 F 11.8
12:00 230.5 F 8.1
14:00 241.9 F 13.1
16:00 241.6 F 12.0
18:00 240.9 F 11.1
20:00 253.9 F 17.8
22:00 265.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.4
2:00 271.8 F 29.2
3:00 272.4 F 29.3
4:00 271.6 F 29.1

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 198.0 F 0.0
10:00 206.5 F 0.0
15:00 210.5 F 0.0
20:00 210.4 F 0.0
25:00 215.2 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:02:04
CYCLE COMPLETE

08/30/2023 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____