

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1913

Name: RITTER2  
Sterilizer ID: V2313228  
01/10/2023 01:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.3 F 0.0  
2:00 150.2 F 2.6  
4:00 174.7 F 6.4  
6:00 197.0 F 11.4  
8:00 216.0 F 5.4  
10:00 230.1 F 10.6  
12:00 231.0 F 7.3  
14:00 241.7 F 12.1  
16:00 244.3 F 13.0  
18:00 257.0 F 19.8  
20:00 268.4 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.5 F 29.7  
2:00 271.5 F 29.3  
3:00 272.5 F 29.8  
4:00 271.7 F 29.2  
  
Min 271.4 F 29.1  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.0 F 0.0  
5:00 233.8 F 0.0  
10:00 233.5 F 0.0  
15:00 234.4 F 0.0  
20:00 216.7 F 0.0  
25:00 214.0 F 0.0  
30:00 210.0 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 21:14  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:32  
CYCLE COMPLETE

01/10/2023 02:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_