
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000643
15:09 15/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:25
136.4 C 224kPa 3:25
136.4 C 225kPa 3:55
136.4 C 225kPa 4:25
136.4 C 225kPa 4:55
136.4 C 225kPa 5:25
136.4 C 225kPa 5:55
136.4 C 224kPa 6:25
136.4 C 224kPa 6:55
136.4 C 224kPa 6:55
VENTING 6:55
AIR DRYING 7:52
CYCLE COMPLETE 8:09

Digital Signature #
E8D10C09AE19FA96
