
Midmark M11 - v1.0.0
Total Cycles: 4358

Name: RITTER 1
Sterilizer ID: V2178198
03/05/2024 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 74.1 F 0.3
4:00 118.7 F 1.2
6:00 111.6 F 3.3
8:00 142.2 F 6.1
10:00 169.2 F 9.4
12:00 188.6 F 13.0
14:00 204.8 F 3.3
16:00 214.7 F 6.4
18:00 225.1 F 9.9
20:00 231.7 F 10.1
22:00 235.4 F 9.7
24:00 238.4 F 9.6
26:00 245.2 F 12.9
28:00 256.3 F 18.8
30:00 266.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.0
2:00 271.5 F 28.3
3:00 272.6 F 29.1
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 222.1 F 0.0
10:00 225.8 F 0.0
15:00 229.4 F 0.0
20:00 226.1 F 0.0
25:00 223.9 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:44
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:10:01
CYCLE COMPLETE

03/05/2024 10:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____