
Midmark M11 - v1.0.0
Total Cycles: 2048

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2022 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.6 F 0.0
2:00 181.3 F 3.6
4:00 186.4 F 8.0
6:00 219.6 F 13.3
8:00 230.5 F 7.5
10:00 240.6 F 12.7
12:00 238.4 F 9.8
14:00 236.4 F 8.4
16:00 249.0 F 14.7
18:00 261.0 F 21.4
20:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.5 F 28.9
2:00 271.4 F 28.1
3:00 271.9 F 28.8
4:00 272.1 F 28.6

Min 271.0 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 223.9 F 0.0
10:00 221.8 F 0.0
15:00 221.6 F 0.0
20:00 217.6 F 0.0
25:00 230.0 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:00
CYCLE COMPLETE

06/22/2022 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____