
Midmark M11 - v1.0.5
Total Cycles: 2920

Name: RITTER2
Sterilizer ID: V2313228
09/25/2023 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 172.2 F 3.1
4:00 189.9 F 7.3
6:00 208.0 F 11.9
8:00 217.2 F 6.7
10:00 232.5 F 11.8
12:00 235.0 F 9.4
14:00 238.4 F 9.6
16:00 247.8 F 14.3
18:00 259.2 F 20.8
20:00 270.0 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.5
1:00 271.9 F 29.4
2:00 271.8 F 29.7
3:00 271.7 F 29.7
4:00 271.8 F 29.8

Min 271.5 F 29.4
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.3 F 0.0
5:00 238.8 F 0.0
10:00 203.5 F 0.2
15:00 196.5 F 0.0
Dry Cycle Aborted

FILLING: 1:39
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:35
DRYING: 18:22
TOTAL CYCLE: 00:49:02

09/25/2023 02:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____