
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001165
9:43 13/07/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 4:14
STERILIZING 5:02
136.0 C 220kPa 5:02
136.2 C 219kPa 5:32
136.4 C 224kPa 6:02
136.3 C 224kPa 6:32
136.3 C 223kPa 7:02
136.5 C 225kPa 7:32
136.1 C 222kPa 8:02
136.4 C 223kPa 8:32
136.3 C 223kPa 9:02
136.3 C 223kPa 9:32
136.4 C 223kPa 10:02
136.4 C 222kPa 10:32
136.4 C 222kPa 11:02
136.3 C 223kPa 11:32
136.2 C 222kPa 12:02
136.1 C 222kPa 12:32
136.4 C 224kPa 13:02
136.2 C 223kPa 13:32
136.3 C 223kPa 14:02
136.2 C 222kPa 14:32
136.2 C 222kPa 15:02
136.2 C 222kPa 15:02
VENTING 15:02
AIR DRYING 15:49
CYCLE COMPLETE 23:11

Digital Signature #
F499FD8595EBD0A4
