
Midmark M11 - v1.0.5
Total Cycles: 3384

Name: RITTER2
Sterilizer ID: V2313228
02/14/2024 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.4 F 0.0
2:00 113.5 F 0.0
4:00 126.9 F 0.4
6:00 152.0 F 3.0
8:00 176.9 F 6.5
10:00 196.1 F 10.2
12:00 211.5 F 7.6
14:00 219.9 F 6.9
16:00 230.9 F 11.1
18:00 231.1 F 7.6
20:00 243.3 F 13.2
22:00 242.6 F 11.9
24:00 254.4 F 18.0
26:00 264.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.4
2:00 271.6 F 29.0
3:00 272.7 F 29.4
4:00 271.6 F 28.8

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 221.6 F 0.0
10:00 228.2 F 0.0
15:00 227.0 F 0.0
20:00 224.2 F 0.0
25:00 221.5 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:57
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:06:18
CYCLE COMPLETE

02/14/2024 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____