

-----  
Midmark M11 - v1.0.0  
Total Cycles: 827

Name: RITTER 1  
Sterilizer ID: V2178198  
05/20/2021 07:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.6 F 0.0  
2:00 86.4 F 0.0  
4:00 84.1 F 0.0  
6:00 122.1 F 0.0  
8:00 151.2 F 0.1  
10:00 178.7 F 0.2  
12:00 199.4 F 1.1  
14:00 216.4 F 4.3  
16:00 224.5 F 7.1  
18:00 232.7 F 10.3  
20:00 233.9 F 10.2  
22:00 238.8 F 10.4  
24:00 237.0 F 8.6  
26:00 246.2 F 12.6  
28:00 245.5 F 12.5  
30:00 257.4 F 18.8  
32:00 267.8 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.8  
1:00 272.5 F 28.4  
2:00 271.5 F 27.7  
3:00 272.6 F 28.5  
4:00 271.7 F 27.7  
  
Min 271.5 F 27.7  
Max 272.7 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 217.6 F 0.0  
10:00 195.0 F 0.0  
15:00 193.5 F 0.0  
20:00 196.0 F 0.0  
25:00 233.8 F 0.0  
30:00 214.3 F 0.0  
  
Drying Complete

FILLING: 0:54  
HEATING: 33:24  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:11:29  
CYCLE COMPLETE

05/20/2021 08:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_