

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4452

Name: RITTER 1  
Sterilizer ID: V2178198  
04/12/2024 11:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.2 F 0.0  
2:00 151.6 F 2.1  
4:00 158.1 F 4.9  
6:00 193.1 F 8.6  
8:00 203.1 F 13.2  
10:00 215.8 F 6.0  
12:00 227.3 F 10.4  
14:00 232.3 F 8.0  
16:00 241.4 F 11.9  
18:00 238.1 F 9.9  
20:00 250.7 F 15.9  
22:00 261.3 F 21.9  
24:00 271.1 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 29.3  
2:00 272.7 F 29.7  
3:00 271.5 F 29.0  
4:00 272.3 F 29.5

Min 271.4 F 29.0  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
5:00 198.1 F 0.0  
10:00 197.1 F 0.0  
15:00 189.2 F 0.0  
20:00 187.3 F 0.0  
25:00 185.5 F 0.0  
30:00 182.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:03  
CYCLE COMPLETE

04/12/2024 12:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_