
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005512
14:35 10/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:00
STERILIZING 3:02
136.4 C 222kPa 3:02
136.1 C 222kPa 3:32
136.4 C 223kPa 4:02
136.4 C 224kPa 4:32
136.4 C 222kPa 5:02
136.3 C 224kPa 5:32
136.3 C 224kPa 6:02
136.3 C 221kPa 6:32
136.3 C 221kPa 6:32
VENTING 6:32
AIR DRYING 7:19
CYCLE COMPLETE 25:27

Digital Signature #
CF21C5EF792F6B6B
