
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002508
9:22 11/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:08
STERILIZING 3:16
136.0 C 222kPa 3:16
136.3 C 223kPa 3:46
136.4 C 224kPa 4:16
136.2 C 224kPa 4:46
136.4 C 224kPa 5:16
136.4 C 225kPa 5:46
136.4 C 224kPa 6:16
136.2 C 223kPa 6:46
136.2 C 223kPa 6:46
VENTING 6:46
AIR DRYING 7:39
CYCLE COMPLETE 19:00

Digital Signature #
0C60A909A1684F0D
