
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
2.2uS / 1.4ppm
CYCLE NUMBER 001869
10:55 18/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:51
STERILIZING 4:03
136.4 C 225kPa 4:03
136.2 C 223kPa 4:33
135.9 C 220kPa 5:03
136.4 C 225kPa 5:33
136.2 C 224kPa 6:03
136.4 C 224kPa 6:33
136.4 C 222kPa 7:03
136.4 C 223kPa 7:33
136.4 C 223kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 8:38

Digital Signature #
1205C15E5B815068
