
Midmark M11 - v1.0.5
Total Cycles: 3205

Name: RITTER2
Sterilizer ID: V2313228
12/11/2023 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.0 F 0.0
2:00 132.8 F 1.0
4:00 159.8 F 4.1
6:00 185.3 F 8.5
8:00 205.9 F 13.2
10:00 226.1 F 7.0
12:00 237.4 F 11.7
14:00 234.6 F 8.5
16:00 241.6 F 11.1
18:00 245.9 F 13.5
20:00 258.2 F 20.2
22:00 269.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.6 F 29.3
2:00 271.7 F 29.2
3:00 272.2 F 29.3
4:00 272.6 F 29.5

Min 271.5 F 29.2
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.8 F 0.0
5:00 229.2 F 0.0
10:00 193.3 F 0.0
15:00 194.1 F 0.0
20:00 192.8 F 0.0
25:00 191.8 F 0.0
30:00 191.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:41
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:15
CYCLE COMPLETE

12/11/2023 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____