

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 004987  
16:19 23/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:31  
STERILIZING 3:18  
136.0 C 220kPa 3:18  
136.4 C 222kPa 3:48  
136.3 C 225kPa 4:18  
136.3 C 222kPa 4:48  
136.4 C 222kPa 5:18  
136.2 C 222kPa 5:48  
136.4 C 224kPa 6:18  
136.0 C 222kPa 6:48  
136.0 C 222kPa 6:48  
VENTING 6:48  
AIR DRYING 7:40  
CYCLE COMPLETE 7:44

-----  
Digital Signature #  
3FAE7D5A66FD76F2  
-----