
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001231
13:48 22/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:37
STERILIZING 4:43
136.5 C 220kPa 4:43
136.3 C 225kPa 5:13
136.2 C 223kPa 5:43
136.2 C 221kPa 6:13
136.4 C 222kPa 6:43
136.2 C 223kPa 7:13
136.2 C 224kPa 7:43
136.4 C 224kPa 8:13
136.4 C 224kPa 8:13
VENTING 8:13
AIR DRYING 9:04
CYCLE COMPLETE 9:06

Digital Signature #
9B4C6EF13ACBEBAB
