
Midmark M11 - v1.0.0
Total Cycles: 3609

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2023 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 71.8 F 0.0
4:00 87.6 F 0.0
6:00 115.9 F 0.0
8:00 144.8 F 0.2
10:00 176.6 F 0.5
12:00 194.3 F 3.9
14:00 207.4 F 6.7
16:00 217.1 F 9.8
18:00 226.8 F 13.3
20:00 229.6 F 7.9
22:00 239.1 F 12.2
24:00 237.1 F 9.5
26:00 238.3 F 9.3
28:00 248.1 F 14.1
30:00 259.6 F 20.5
32:00 270.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.3
3:00 272.5 F 28.7
4:00 271.4 F 28.1

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 198.1 F 0.0
10:00 199.2 F 0.0
15:00 205.7 F 0.0
20:00 211.1 F 0.0
25:00 215.4 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 32:56
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:10:59
CYCLE COMPLETE

08/03/2023 09:31 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____