
Midmark M11 - v1.0.5
Total Cycles: 2281

Name: RITTER2
Sterilizer ID: V2313228
04/21/2023 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 182.9 F 2.5
4:00 190.7 F 7.0
6:00 210.9 F 12.6
8:00 227.7 F 8.3
10:00 236.5 F 11.6
12:00 240.7 F 12.0
14:00 240.6 F 11.1
16:00 254.2 F 18.0
18:00 265.8 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.5
2:00 271.3 F 28.9
3:00 272.9 F 29.6
4:00 271.7 F 29.0

Min 271.2 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 217.9 F 0.0
10:00 232.5 F 0.0
15:00 228.5 F 0.0
20:00 225.7 F 0.0
25:00 225.5 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 19:38
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

04/21/2023 10:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____