
Midmark M11 - v1.0.0
Total Cycles: 2120

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2022 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.0 F 0.0
2:00 169.1 F 2.1
4:00 181.4 F 5.9
6:00 207.5 F 10.3
8:00 220.2 F 5.3
10:00 230.0 F 9.6
12:00 236.7 F 9.6
14:00 241.5 F 11.8
16:00 238.6 F 9.9
18:00 251.9 F 16.3
20:00 262.7 F 22.7
22:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.0 F 28.7
2:00 272.4 F 29.2
3:00 271.6 F 28.7
4:00 271.8 F 28.7

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 237.3 F 0.0
10:00 229.7 F 0.0
15:00 229.1 F 0.0
20:00 229.8 F 0.0
25:00 230.2 F 0.0
30:00 232.2 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

07/13/2022 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____