
Midmark M11 - v1.0.5
Total Cycles: 2701

Name: RITTER2
Sterilizer ID: V2313228
07/25/2023 05:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 184.5 F 3.6
4:00 204.2 F 8.9
6:00 219.4 F 8.9
8:00 231.6 F 9.7
10:00 233.6 F 8.8
12:00 243.8 F 12.7
14:00 244.9 F 13.1
16:00 257.6 F 19.9
18:00 268.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.5 F 29.3
2:00 272.2 F 29.7
3:00 272.1 F 29.0
4:00 272.6 F 29.7

Min 271.6 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 214.9 F 0.0
10:00 228.8 F 0.0
15:00 229.5 F 0.0
20:00 228.0 F 0.0
25:00 226.7 F 0.0
30:00 224.7 F 0.0

Drying Complete

FILLING: 2:31
HEATING: 19:06
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

07/25/2023 06:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____