

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2023

Name: RITTER 1  
Sterilizer ID: V2178198  
06/15/2022 08:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.2 F 0.0  
2:00 154.1 F 1.0  
4:00 166.2 F 3.7  
6:00 186.3 F 7.3  
8:00 210.4 F 12.5  
10:00 223.5 F 6.5  
12:00 235.1 F 11.4  
14:00 232.9 F 7.8  
16:00 244.7 F 13.2  
18:00 243.5 F 11.8  
20:00 255.6 F 18.2  
22:00 266.6 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.2  
1:00 272.2 F 29.0  
2:00 271.7 F 28.2  
3:00 272.3 F 29.0  
4:00 271.7 F 28.3

Min 271.3 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 214.9 F 0.0  
10:00 230.7 F 0.0  
15:00 230.2 F 0.0  
20:00 228.4 F 0.0  
25:00 237.8 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 23:28  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:25  
CYCLE COMPLETE

06/15/2022 09:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_