
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000144
6:37 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 3:00
STERILIZING 4:01
136.0 C 220kPa 4:01
136.3 C 221kPa 4:31
136.3 C 221kPa 5:01
136.4 C 222kPa 5:31
136.3 C 223kPa 6:01
136.3 C 222kPa 6:31
136.4 C 222kPa 7:01
135.9 C 219kPa 7:31
135.9 C 219kPa 7:31
VENTING 7:32
AIR DRYING 8:22
CYCLE COMPLETE 8:26

Digital Signature #
662F572ACDDD8C50
