

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1866

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2022 04:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
2:00 192.9 F 4.1  
4:00 216.5 F 9.1  
6:00 221.5 F 7.5  
8:00 227.0 F 10.5  
10:00 231.8 F 8.0  
12:00 240.7 F 10.9  
14:00 247.5 F 14.2  
16:00 260.4 F 21.5  
18:00 271.9 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.2 F 29.3  
2:00 271.5 F 29.3  
3:00 272.1 F 29.2  
4:00 271.3 F 29.0  
  
Min 271.2 F 28.7  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.9 F 0.0  
5:00 212.2 F 0.0  
10:00 225.1 F 0.0  
15:00 229.0 F 0.0  
20:00 225.8 F 0.0  
25:00 224.4 F 0.0  
30:00 222.8 F 0.0  
  
Drying Complete

FILLING: 2:03  
HEATING: 18:31  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:43  
CYCLE COMPLETE

12/20/2022 05:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_