
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001358
11:00 23/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:25
STERILIZING 4:58
136.4 C 225kPa 4:58
136.3 C 225kPa 5:29
136.4 C 224kPa 5:59
136.2 C 222kPa 6:29
136.4 C 224kPa 6:59
136.4 C 225kPa 7:29
136.4 C 226kPa 7:59
136.4 C 226kPa 8:29
136.4 C 226kPa 8:29
VENTING 8:29
AIR DRYING 9:28
CYCLE COMPLETE 24:28

Digital Signature #
4DFCBB9674F36EFE
