
Midmark M11 - v1.0.0
Total Cycles: 851

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2021 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 172.5 F 2.1
4:00 186.1 F 6.2
6:00 211.2 F 11.4
8:00 222.7 F 6.1
10:00 236.2 F 11.5
12:00 235.1 F 8.7
14:00 240.8 F 10.2
16:00 247.7 F 13.8
18:00 260.2 F 20.6
20:00 271.3 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.4 F 28.3
2:00 271.9 F 28.4
3:00 272.0 F 28.1
4:00 272.2 F 28.6

Min 271.2 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 230.4 F 0.0
10:00 245.8 F 0.0
15:00 221.0 F 0.0
20:00 205.0 F 0.0
25:00 187.4 F 0.0
30:00 174.8 F 0.2
Drying Complete

FILLING: 1:38
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

05/31/2021 11:50 AM
ACCEPTED: X
REJECTED:
Operator: U01