

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4183

Name: RITTER2  
Sterilizer ID: V2313228  
12/04/2024 02:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
2:00 183.9 F 1.1  
4:00 189.6 F 5.4  
6:00 209.7 F 9.6  
8:00 222.8 F 9.6  
10:00 230.3 F 9.5  
12:00 238.3 F 9.8  
14:00 243.0 F 13.0  
16:00 243.1 F 12.3  
18:00 255.5 F 18.9  
20:00 266.7 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.6 F 29.7  
2:00 271.5 F 29.0  
3:00 272.7 F 29.9  
4:00 272.0 F 29.3  
  
Min 271.4 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 214.2 F 0.0  
10:00 224.5 F 0.0  
15:00 224.1 F 0.0  
20:00 219.5 F 0.0  
25:00 217.1 F 0.0  
30:00 211.9 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:33  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:01:03  
CYCLE COMPLETE

12/04/2024 03:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_