
Midmark M11 - v1.0.0
Total Cycles: 4407

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2024 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.0 F 0.0
2:00 113.4 F 1.0
4:00 130.0 F 3.2
6:00 159.5 F 6.7
8:00 181.2 F 10.8
10:00 204.6 F 2.6
12:00 222.3 F 6.4
14:00 233.2 F 10.4
16:00 235.6 F 7.9
18:00 240.6 F 11.5
20:00 237.6 F 9.2
22:00 250.6 F 15.4
24:00 261.7 F 21.7
26:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.7
2:00 271.5 F 28.2
3:00 272.7 F 28.8
4:00 271.5 F 28.0

Min 271.4 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 204.6 F 0.0
10:00 214.1 F 0.0
15:00 222.2 F 0.0
20:00 226.8 F 0.0
25:00 227.1 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:32
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:56
CYCLE COMPLETE

03/26/2024 04:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____