
Midmark M11 - v1.0.5
Total Cycles: 1245

Name: RITTER2
Sterilizer ID: V2313228
06/29/2022 04:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.9 F 0.0
2:00 181.4 F 3.6
4:00 193.0 F 8.4
6:00 219.5 F 9.8
8:00 227.4 F 9.9
10:00 233.0 F 8.4
12:00 242.3 F 12.3
14:00 245.4 F 13.3
16:00 257.9 F 20.2
18:00 269.3 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.7
2:00 271.4 F 29.1
3:00 272.4 F 29.8
4:00 271.7 F 29.1

Min 271.3 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 225.5 F 0.0
10:00 244.2 F 0.0
15:00 227.2 F 0.0
20:00 219.1 F 0.0
25:00 217.7 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

06/29/2022 05:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____