
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001575
13:54 08/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:21
STERILIZING 3:19
136.0 C 221kPa 3:19
136.4 C 226kPa 3:49
136.4 C 226kPa 4:19
136.4 C 225kPa 4:49
136.1 C 223kPa 5:19
136.2 C 223kPa 5:49
136.4 C 226kPa 6:19
136.3 C 225kPa 6:49
136.3 C 225kPa 6:49
VENTING 6:50
AIR DRYING 7:35
CYCLE COMPLETE 8:00

Digital Signature #
009F2C54A3E34974
