
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004119
10:05 27/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:22
STERILIZING 3:17
136.0 C 220kPa 3:17
136.3 C 224kPa 3:47
136.1 C 217kPa 4:17
136.3 C 223kPa 4:47
136.4 C 221kPa 5:17
136.3 C 221kPa 5:47
136.3 C 221kPa 6:17
136.4 C 223kPa 6:47
136.4 C 223kPa 6:47
VENTING 6:47
AIR DRYING 7:37
CYCLE COMPLETE 7:55

Digital Signature #
514A771B72856947
