
Midmark M11 - v1.0.5
Total Cycles: 1298

Name: RITTER2
Sterilizer ID: V2313228
07/14/2022 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 167.4 F 0.4
4:00 195.3 F 4.9
6:00 212.4 F 9.7
8:00 222.9 F 8.8
10:00 233.8 F 10.8
12:00 233.5 F 8.5
14:00 241.0 F 10.5
16:00 247.1 F 14.3
18:00 259.5 F 21.1
20:00 270.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.5
2:00 271.9 F 29.3
3:00 272.3 F 29.3
4:00 271.9 F 29.4

Min 271.6 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 211.9 F 0.0
10:00 207.0 F 0.0
15:00 195.9 F 0.0
20:00 186.8 F 0.0
25:00 180.7 F 0.0
30:00 176.0 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

07/14/2022 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____