

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002495  
7:57 22/10/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:31  
STERILIZING 3:37  
136.5 C 223kPa 3:37  
136.2 C 221kPa 4:07  
136.4 C 219kPa 4:37  
136.1 C 220kPa 5:07  
136.4 C 223kPa 5:37  
136.2 C 217kPa 6:07  
136.4 C 223kPa 6:37  
136.3 C 222kPa 7:07  
136.3 C 222kPa 7:07  
VENTING 7:07  
AIR DRYING 7:58  
CYCLE COMPLETE 8:01  
-----

Digital Signature #  
40EE76D196D5E11B  
-----