
Midmark M11 - v1.0.5
Total Cycles: 3706

Name: RITTER2
Sterilizer ID: V2313228
06/17/2024 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.2 F 0.0
2:00 138.4 F 1.8
4:00 145.4 F 4.6
6:00 178.7 F 8.4
8:00 192.8 F 13.0
10:00 216.0 F 6.7
12:00 229.1 F 11.3
14:00 230.4 F 7.9
16:00 242.8 F 13.5
18:00 242.0 F 11.9
20:00 253.8 F 18.1
22:00 264.6 F 24.5
24:00 271.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 271.9 F 29.5
2:00 272.4 F 29.7
3:00 272.4 F 30.0
4:00 272.2 F 30.2

Min 271.5 F 29.5
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 209.6 F 0.0
10:00 192.8 F 0.0
15:00 194.2 F 0.0
20:00 196.5 F 0.0
25:00 205.6 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:10
CYCLE COMPLETE

06/17/2024 12:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____