
Midmark M11 - v1.0.0
Total Cycles: 3004

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2023 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 190.6 F 3.8
4:00 212.8 F 9.3
6:00 224.5 F 6.9
8:00 236.4 F 11.2
10:00 235.4 F 9.0
12:00 238.6 F 9.3
14:00 249.6 F 15.2
16:00 262.5 F 22.5
18:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.6 F 28.7
3:00 272.4 F 28.9
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 217.3 F 0.0
10:00 225.0 F 0.0
15:00 225.5 F 0.0
20:00 229.8 F 0.0
25:00 227.9 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:22
CYCLE COMPLETE

03/14/2023 01:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____