
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002427
10:06 17/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:09
STERILIZING 3:12
136.0 C 220kPa 3:12
136.0 C 222kPa 3:42
136.4 C 221kPa 4:12
135.9 C 221kPa 4:42
136.4 C 223kPa 5:12
136.4 C 224kPa 5:42
136.2 C 223kPa 6:12
136.4 C 224kPa 6:42
136.4 C 224kPa 6:42
VENTING 6:42
AIR DRYING 7:33
CYCLE COMPLETE 19:47

Digital Signature #
83E01D564065CA3B
