

-----  
Midmark M11 - v1.0.5  
Total Cycles: 215

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2021 10:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.7 F 0.0  
2:00 192.0 F 4.5  
4:00 216.0 F 10.7  
6:00 228.6 F 7.3  
8:00 240.5 F 13.3  
10:00 242.4 F 12.1  
12:00 242.5 F 11.7  
14:00 256.2 F 18.8  
16:00 267.9 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.2 F 29.2  
2:00 271.5 F 29.1  
3:00 272.1 F 29.1  
4:00 271.2 F 28.8  
  
Min 271.2 F 28.6  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 243.3 F 0.0  
10:00 223.2 F 0.0  
15:00 226.9 F 0.0  
20:00 222.1 F 0.0  
25:00 232.8 F 0.0  
30:00 229.6 F 0.0  
  
Drying Complete

FILLING: 2:26  
HEATING: 17:15  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:56:53  
CYCLE COMPLETE

07/15/2021 11:59 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02