
Midmark M11 - v1.0.0
Total Cycles: 1057

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2021 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 172.3 F 1.4
4:00 172.8 F 4.5
6:00 194.1 F 8.6
8:00 214.0 F 10.5
10:00 224.3 F 7.8
12:00 237.9 F 12.8
14:00 236.5 F 9.5
16:00 238.2 F 9.2
18:00 247.3 F 13.8
20:00 259.7 F 20.6
22:00 270.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.8
2:00 271.4 F 28.3
3:00 272.4 F 28.9
4:00 271.5 F 28.2

Min 271.3 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 212.9 F 0.0
10:00 207.3 F 0.0
15:00 203.1 F 0.0
20:00 185.6 F 0.0
25:00 182.3 F 0.0
30:00 192.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:05
CYCLE COMPLETE

07/27/2021 08:57 AM
ACCEPTED: X
REJECTED:
Operator: U01