
Midmark M11 - v1.0.5
Total Cycles: 3891

Name: RITTER2
Sterilizer ID: V2313228
08/26/2024 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.4 F 0.0
2:00 67.8 F 0.0
4:00 75.2 F 0.0
6:00 106.6 F 1.0
8:00 139.5 F 3.4
10:00 158.5 F 6.0
12:00 180.7 F 8.9
14:00 198.5 F 12.1
16:00 211.0 F 4.6
18:00 217.9 F 8.0
20:00 228.4 F 11.7
22:00 228.4 F 7.2
24:00 240.3 F 12.0
26:00 237.7 F 9.8
28:00 249.4 F 15.4
30:00 259.6 F 21.1
32:00 269.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.6
2:00 271.4 F 28.9
3:00 272.7 F 29.7
4:00 271.6 F 28.9

Min 271.3 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 216.0 F 0.0
10:00 185.3 F 0.0
15:00 181.9 F 0.0
20:00 177.3 F 0.0
25:00 171.4 F 0.0
30:00 166.4 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:10
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:11:29
CYCLE COMPLETE

08/26/2024 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____