
Midmark M11 - v1.0.0
Total Cycles: 3467

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2023 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 195.4 F 3.7
4:00 203.3 F 8.8
6:00 222.8 F 7.7
8:00 233.2 F 10.0
10:00 237.9 F 8.3
12:00 244.4 F 12.5
14:00 245.2 F 12.8
16:00 258.1 F 19.7
18:00 269.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.4 F 29.1
2:00 272.1 F 28.5
3:00 271.9 F 28.7
4:00 272.5 F 28.9

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 217.6 F 0.0
10:00 233.3 F 0.0
15:00 231.0 F 0.0
20:00 225.5 F 0.0
25:00 224.6 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

06/29/2023 01:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____