
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001402
14:39 09/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:33
STERILIZING 3:40
136.0 C 220kPa 3:40
136.3 C 222kPa 4:10
136.3 C 222kPa 4:40
136.2 C 224kPa 5:10
136.2 C 222kPa 5:40
136.4 C 225kPa 6:10
136.4 C 225kPa 6:40
136.4 C 223kPa 7:10
136.4 C 223kPa 7:10
VENTING 7:10
AIR DRYING 8:01
CYCLE COMPLETE 12:58

Digital Signature #
E386E1620CC1D5D2
