

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4490

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2025 03:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
2:00 177.3 F 3.1  
4:00 205.0 F 8.8  
6:00 221.0 F 7.9  
8:00 234.0 F 10.2  
10:00 234.2 F 8.0  
12:00 241.0 F 10.5  
14:00 248.5 F 14.6  
16:00 261.6 F 22.0  
18:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 272.2 F 28.7  
2:00 272.7 F 29.2  
3:00 272.1 F 29.0  
4:00 272.1 F 28.7  
  
Min 271.6 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.8 F 0.0  
5:00 227.6 F 0.0  
10:00 223.5 F 0.0  
15:00 218.2 F 0.0  
20:00 206.3 F 0.0  
25:00 205.0 F 0.0  
30:00 201.9 F 0.0  
Drying Complete

FILLING: 2:17  
HEATING: 18:14  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:57:49  
CYCLE COMPLETE

04/04/2025 04:42 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_