
Midmark M11 - v1.0.5
Total Cycles: 3381

Name: RITTER2
Sterilizer ID: V2313228
02/13/2024 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.0 F 0.0
2:00 118.1 F 0.4
4:00 130.4 F 2.4
6:00 159.2 F 5.9
8:00 183.3 F 10.1
10:00 207.0 F 5.7
12:00 220.3 F 8.4
14:00 233.4 F 13.4
16:00 235.6 F 10.5
18:00 234.8 F 8.8
20:00 248.9 F 15.6
22:00 261.1 F 22.4
24:00 272.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.3 F 29.6
2:00 272.3 F 30.1
3:00 271.8 F 29.3
4:00 272.6 F 30.0

Min 271.4 F 29.3
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 211.1 F 0.0
10:00 199.5 F 0.0
15:00 197.7 F 0.0
20:00 195.8 F 0.0
25:00 194.2 F 0.0
30:00 194.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:29
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

02/13/2024 01:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____