

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4169

Name: RITTER 1  
Sterilizer ID: V2178198  
12/29/2023 11:22 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.2 F 0.0  
2:00 156.6 F 1.7  
4:00 160.6 F 5.0  
6:00 187.4 F 10.4  
8:00 213.8 F 4.3  
10:00 233.2 F 10.8  
12:00 233.3 F 8.3  
14:00 238.9 F 9.7  
16:00 247.7 F 14.0  
18:00 261.0 F 21.4  
20:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.2  
1:00 272.4 F 28.6  
2:00 272.7 F 29.0  
3:00 272.1 F 28.8  
  
Min 271.4 F 28.0  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 227.7 F 0.0  
5:00 224.8 F 0.0  
Dry Cycle Aborted

FILLING: 1:26  
HEATING: 20:09  
STERILIZING: 3:00  
VENTING: 2:60  
DRYING: 5:43  
TOTAL CYCLE: 00:33:18

12/29/2023 11:55 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_