
Midmark M11 - v1.0.5
Total Cycles: 2883

Name: RITTER2
Sterilizer ID: V2313228
09/14/2023 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 173.3 F 3.4
4:00 189.4 F 7.6
6:00 206.1 F 12.4
8:00 224.2 F 7.4
10:00 235.3 F 12.6
12:00 235.9 F 10.2
14:00 236.5 F 8.8
16:00 249.4 F 15.6
18:00 261.1 F 22.2
20:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.8 F 29.9
2:00 272.0 F 29.4
3:00 271.0 F 29.2
4:00 272.7 F 30.2

Min 271.0 F 29.0
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 200.2 F 0.0
10:00 200.3 F 0.0
15:00 201.0 F 0.0
20:00 203.6 F 0.0
25:00 207.3 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

09/14/2023 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____