
Midmark M11 - v1.0.0
Total Cycles: 1218

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2021 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.7 F 0.0
2:00 197.6 F 3.9
4:00 201.5 F 8.7
6:00 228.3 F 8.6
8:00 232.1 F 10.3
10:00 233.4 F 7.5
12:00 241.1 F 10.9
14:00 247.7 F 13.7
16:00 260.4 F 20.8
18:00 271.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.2 F 28.6
2:00 272.7 F 28.9
3:00 272.3 F 29.1
4:00 271.6 F 28.5

Min 271.5 F 28.4
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 245.4 F 0.0
10:00 236.6 F 0.0
15:00 227.9 F 0.0
20:00 229.6 F 0.0
25:00 228.2 F 0.0
30:00 225.9 F 0.0

Drying Complete

FILLING: 1:45
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

09/29/2021 11:17 AM
ACCEPTED: X
REJECTED:
Operator: U01