

-----  
Midmark M11 - v1.0.0  
Total Cycles: 650

Name: MIDMARK  
Sterilizer ID: V2178198  
02/02/2021 01:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.8 F 0.0  
2:00 114.9 F 0.1  
4:00 130.0 F 0.8  
6:00 158.6 F 3.7  
8:00 192.9 F 7.7  
10:00 207.1 F 12.7  
12:00 222.8 F 6.6  
14:00 234.6 F 11.6  
16:00 235.5 F 8.5  
18:00 240.2 F 10.5  
20:00 245.9 F 13.0  
22:00 258.2 F 19.5  
24:00 269.3 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.0 F 28.9  
2:00 272.4 F 28.9  
3:00 272.4 F 28.8  
4:00 272.4 F 28.7

Min 271.7 F 28.4  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.1 F 0.0  
5:00 225.5 F 0.0  
10:00 205.3 F 0.0  
15:00 202.8 F 0.0  
20:00 203.3 F 0.0  
25:00 208.9 F 0.0  
30:00 215.0 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:33  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:03:35  
CYCLE COMPLETE

02/02/2021 02:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_