

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4270

Name: RITTER2  
Sterilizer ID: V2313228  
01/16/2025 12:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 106.9 F 0.0  
2:00 106.8 F 0.0  
4:00 113.0 F 0.0  
6:00 148.5 F 0.6  
8:00 170.8 F 2.1  
10:00 191.8 F 5.4  
12:00 207.9 F 9.0  
14:00 219.8 F 13.1  
16:00 229.5 F 8.4  
18:00 240.5 F 13.2  
20:00 239.7 F 11.1  
22:00 237.3 F 9.7  
24:00 250.9 F 16.1  
26:00 262.0 F 22.6  
28:00 272.3 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.8 F 29.9  
2:00 272.8 F 30.0  
3:00 272.5 F 30.0  
4:00 272.2 F 29.7

Min 271.0 F 28.8  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.3 F 0.0  
5:00 208.4 F 0.0  
10:00 213.2 F 0.0  
15:00 225.1 F 0.0  
20:00 226.1 F 0.0  
25:00 225.8 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 28:23  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:07:09  
CYCLE COMPLETE

01/16/2025 01:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_