
Midmark M11 - v1.0.0
Total Cycles: 2810

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 188.0 F 0.0
4:00 185.7 F 2.7
6:00 209.9 F 7.5
8:00 223.9 F 12.8
10:00 228.1 F 8.3
12:00 240.1 F 12.5
14:00 241.0 F 11.4
16:00 238.9 F 9.9
18:00 252.8 F 16.8
20:00 264.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.5 F 28.2
3:00 272.7 F 29.1
4:00 272.0 F 28.5

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 213.0 F 0.0
10:00 231.5 F 0.0
15:00 228.5 F 0.0
20:00 228.4 F 0.0
25:00 229.9 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

01/24/2023 10:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____