
Midmark M11 - v1.0.0
Total Cycles: 2838

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2023 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.5 F 0.0
2:00 77.9 F 0.0
4:00 90.9 F 0.1
6:00 127.0 F 1.3
8:00 157.2 F 4.2
10:00 178.6 F 7.5
12:00 200.1 F 11.1
14:00 210.2 F 4.5
16:00 219.1 F 6.6
18:00 229.3 F 10.3
20:00 232.1 F 8.5
22:00 238.8 F 10.8
24:00 234.3 F 7.9
26:00 248.6 F 14.6
28:00 260.1 F 21.0
30:00 271.0 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.1
2:00 271.8 F 28.3
3:00 271.8 F 28.8
4:00 272.5 F 28.9

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 217.3 F 0.0
10:00 225.3 F 0.0
15:00 233.9 F 0.0
20:00 214.9 F 0.0
25:00 208.0 F 0.0
30:00 206.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:46
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

02/01/2023 09:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____