
Midmark M11 - v1.0.5
Total Cycles: 3371

Name: RITTER2
Sterilizer ID: V2313228
02/09/2024 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 75.9 F 0.0
4:00 84.5 F 0.0
6:00 102.5 F 0.9
8:00 145.5 F 3.7
10:00 171.2 F 7.2
12:00 189.1 F 11.1
14:00 210.9 F 4.2
16:00 220.1 F 7.3
18:00 230.6 F 11.3
20:00 229.8 F 7.4
22:00 242.0 F 12.8
24:00 240.5 F 11.2
26:00 252.9 F 17.3
28:00 263.9 F 23.7
30:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.2 F 29.3
2:00 271.9 F 29.6
3:00 271.7 F 29.0
4:00 272.3 F 29.8

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 209.3 F 0.0
10:00 197.9 F 0.0
15:00 194.2 F 0.0
20:00 189.9 F 0.0
25:00 187.1 F 0.0
30:00 185.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:09
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:12
CYCLE COMPLETE

02/09/2024 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____